



Contribution ID: 75

Type: Poster

Applicability of k_{00} -based neutron activation analysis using a Compton suppression & gamma-ray spectrometer

Tuesday, April 20, 2010 11:45 AM (20 minutes)

Compton suppression system (CSS) in conjunction with γ -ray spectrometer allows to lower the detection limits for certain elements by reducing the background levels in the acquired γ -ray spectra. The calibration of CSS for use in the k_{00} -based neutron activation analysis (k_{00} -NAA) has been performed in normal mode and applied to Compton suppression mode. In this case, efficiency data are only applied to radionuclides that emit non-coincident γ -rays. In this work the applicability of the k_{00} -NAA method using CSS (k_{00} -CSNAA) in Compton suppression mode for radionuclides that emit coincident γ -rays is tested. The SMELS (a synthetic multi-element standard) consisting of nuclides with different γ -ray energy and Q_{00} values was used to evaluate the Compton suppression factor for several radionuclides of interest, e.g. ^{60}Co , ^{51}Cr , ^{152}Eu , ^{59}Fe , ^{75}Se , ^{46}Sc and ^{85}Sr . The applicability of k_{00} -CSNAA has been tested by analyzing several reference materials, i.e. Coal Fly Ash (NIST-SRM-1633a), Soil (IAEA-Soil-7) and Lichen Material (IAEA-CRM-336). Preliminary results of the test are presented and discussed.

Primary author: Dr HO, Dung Manh (Technological and Nuclear Institute, Sacavem, Portugal)

Co-authors: Dr BEASLEY, Daniel (Technological and Nuclear Institute, Sacavem, Portugal); Dr FREITAS, Maria do Carmo (Technological and Nuclear Institute, Sacavem, Portugal); Mr CANHA, Nuno (Technological and Nuclear Institute, Sacavem, Portugal)

Presenter: Dr HO, Dung Manh (Technological and Nuclear Institute, Sacavem, Portugal)

Session Classification: Poster Session - Nuclear Analytical Methods

Track Classification: Nuclear Analytical Methods