



Contribution ID: 897

Type: Verbal

## Development of new low-level method for the analysis of iodine 129 and the isotopic ratio $^{129}\text{I}/^{127}\text{I}$ determination

*Tuesday, 17 May 2022 14:00 (20 minutes)*

Iodine is an extremely volatile element naturally present under various organic or inorganic forms. These physico-chemical characteristics complicate its extraction and quantification. Among the 37 iodine isotopes,  $^{127}\text{I}$  is the only stable one and  $^{129}\text{I}$  is the radioisotope with the longest half-life ( $16.1 \times 10^6$  y).  $^{129}\text{I}$  is chronically and regulatory released by fuel reprocessing plants e.g. Sellafield (UK) and La Hague (France) and also released in the atmosphere by nuclear power plants during normal operation process or in case of an accident.  $^{129}\text{I}$  quantification and the determination of  $^{129}\text{I}/^{127}\text{I}$  isotopic ratio is then essential in a post-accidental context, environmental monitoring field or transfer mechanism studies.

$^{129}\text{I}$  quantification is usually performed by  $\gamma$ -spectroscopy or Liquid Scintillation Counting (LSC). These methods don't however allow the determination of the isotopic ratio ( $^{129}\text{I}/^{127}\text{I}$ ) and obtained limit of detection (20 mBq/L and 10 mBq/L respectively after chemical treatment) are often higher than the environmental background. AMS allows also the quantification of  $^{129}\text{I}$  and  $^{129}\text{I}/^{127}\text{I}$ . Although performances of AMS no longer have to be demonstrated (excellent sensibility and selectivity), this device is not easily accessible.

ICP-MS is an excellent alternative for the determination of  $^{129}\text{I}$  and  $^{129}\text{I}/^{127}\text{I}$  isotopic ratio. However, spectral and non-spectral interferences could disturb the measurement. In the present work, the ICP-MS measurements were performed using an ICP-MS/MS (8900 Agilent®). The measurement medium was meticulously studied to minimize non-spectral interferences such as memory and matrix effects. Concerning spectral interferences, the octopole collision/reaction cell and the two quadrupole mass filters allowed minimizing polyatomic interferences. The isobaric interference due to  $^{129}\text{Xe}$  was eliminated after studying different reaction gas ( $\text{O}_2$ ,  $\text{N}_2\text{O}$  and  $\text{CO}_2$ ) with on-mass and mass-shift detection modes.

The new method allows the measurement of iodine 129 either directly or after chemical treatment. The chemical treatment was developed, adapted to ICP-MS measurement medium and validated with reference materials. The turnaround time for the experiment is less than a day. A gain in sensitivity up to a factor 200 was observed and allowed reducing the detection limit up to 0.06 mBq/L after chemical treatment, which is 100 times lower than some current methods.

**Primary author:** Ms CARRIER, Coralie (Institut de Radioprotection et de Sûreté Nucléaire (IRSN), PSE-ENV/SAME/LERCA)

**Co-authors:** Dr HABIBI, Azza (Institut de Radioprotection et de Sûreté Nucléaire (IRSN), PSE-ENV/SAME/LERCA); Mrs AUGERAY, Céline (Institut de Radioprotection et de Sûreté Nucléaire (IRSN), PSE-ENV/SAME/LERCA); Dr MARO, Denis (Institut de Radioprotection et de Sûreté Nucléaire (IRSN), PSE-ENV/SRTE/LRC); Dr BENEDETTI, Lucilla (Univ. Aix-Marseille, CNRS, IRD, INRAE, Coll. France, UM 34 CEREGE)

**Presenter:** Ms CARRIER, Coralie (Institut de Radioprotection et de Sûreté Nucléaire (IRSN), PSE-ENV/SAME/LERCA)

**Session Classification:** Nuclear Analytical Methods

**Track Classification:** Nuclear Analytical Methods