



Contribution ID: 82

Type: Verbal

## A comparative study to determine $^{63}\text{Ni}$ in Low and Intermediate Level Nuclear Wastes

Friday, 16 May 2014 10:45 (15 minutes)

The LASE (Operator Support Analyses Laboratory) develops and implements radiochemical methods devoted to the characterization of Long-Lived Radionuclides in Low and Intermediate Level Nuclear Wastes.  $^{63}\text{Ni}$  is generally one of the radionuclides which impose the classification of the radioactive wastes. This radionuclide is produced by neutron activation of stable nickel which is a major component of construction materials used in nuclear industry. As a pure beta emitter,  $^{63}\text{Ni}$  must be isolated from the matrices and the interfering radionuclides through chemical separations prior to any measurement [1]. Due to its half-life (98.7 years), the French National Radioactive Waste Management Agency (ANDRA) requests its monitoring in waste packages from an activity threshold of 1 Bq/g [2]. Consequently, a rapid and selective radiochemical procedure is needed to measure  $^{63}\text{Ni}$  in nuclear wastes.

The reference radiochemical method which has been used for a long time in the laboratory is standardized [3]. It is based on the extraction of a Ni-dimethylglyoxime (Ni-DMG2) complex with chloroform followed by a back-extraction with diluted HCl. The aim of this work is to determine an alternative protocol that eliminates chloroform which is restricted through REACH regulation. An alternative method can be the precipitation of Ni-DMG2 complex followed by its dissolution in  $\text{HNO}_3$  [3]. Another procedure consist in implementing extraction chromatography where Ni precipitates on Ni<sup>®</sup> resin impregnated with DMG and is then eluted in  $\text{HNO}_3$  [4,5].

The presentation will first focus on the optimization of the radiochemical procedures by performing speciation calculations with Chess<sup>®</sup> software. The second part of the work will be addressed to the application of the three separation methods to various types of radioactive wastes (polymers, ion exchange resins, effluents and steels). For all procedures,  $^{63}\text{Ni}$  is analyzed by liquid scintillation counting (LSC) whereas the separation yield is determined from stable Ni measured by ICP-AES. The results will be compared in terms of accuracy, separation yields, analysis time and waste production.

### References:

- [1] X. Hou et al., Anal. Chim. Acta 608 (2008) 105-139.
- [2] ANDRA (2008) ACO.SP.ASRE.99.002 acceptance specifications for nuclear wastes.
- [3] AFNOR (2001) NF M 60-317 standard.
- [4] O. Rosskopfova et al., J. Radioanal. Nucl. Chem. 289 (2011) 251-256.
- [5] P.E. Warwick et al., Anal. Chim. Acta 567 (2006) 277-285.

**Primary author:** Dr GAUTIER, Celine (CEA)

**Co-authors:** GARCIA, Cecile (CEA); COLIN, Christele (CEA); GOUTELARD, Florence (CEA)

**Presenter:** Dr GAUTIER, Celine (CEA)

**Session Classification:** Separation Methods, Speciation 5

**Track Classification:** Separation Methods, Speciation