



Contribution ID: 52

Type: Poster

Semi-automated Procedure for Rapid Determination of ^{90}Sr in Water Samples by Cherenkov Counting

Tuesday, 13 May 2014 17:15 (1h 30m)

Radiostrontium as a high yield fission product with long physical (28.8 years) and biological half-life (~49 years), is one of the most hazardous radiocontaminants in the environment. Therefore, almost all environmental radioactivity monitoring programs include its quantitative determination. Due to its radiochemical properties standard procedure for its determination is complicated and time consuming. Consequently, so called rapid methods for its separation and determination are being rapidly developed [1-4]. However, there is a lack of prompt and reliable methods for determination of low level activities in the environmental samples. Hence, in this paper, method for rapid low level activity determination which includes simultaneous selective binding of ^{90}Sr and ^{90}Y on chromatographic column filed with mixture of DGA-SuperLig resins and subsequent on column Cherenkov detection will be presented. Highly selective strontium binding on SuperLig 620 and yttrium on DGA resin are based on molecular recognition [5,6] and enable their selective and quantitative binding with almost 100% recovery. If one assume that in the sample ^{90}Sr is in equilibrium with its daughter ^{90}Y , mixture of SuperLig 620 and DGA resin might enable rapid selective separation of ^{90}Sr and ^{90}Y from matrices and direct on column Cherenkov counting. Therefore, method consist of automated sample delivery to the column filled with DGA/SuperLig resins and ^{90}Sr - ^{90}Y Cherenkov counting on low level counter, TriCarb 3180 TR/SL. Sample is delivered to the column at constant flow rate until the breakthrough point and ^{90}Sr via ^{90}Y was determined by counting column in PE vial surrounded with 5 M HNO_3 to achieve higher efficiency determination. Thereby, to develop this method, best mixing options using different media were examined as well as breakthrough curves for strontium and yttrium in 0.5M HNO_3 were determined. The new method significantly reduces time and human labor for determination of radioactive strontium in water samples. It was validated with proficiency testing water samples as well as natural water samples.

1. Fiskum SK, Riley RG, Thompson CJ (2000) J Radioanal Nucl Chem 245(2):261-272
2. O'Hara MJ, Burge SR, Grate JW (2009) Anal Chem 81:1228-1237
3. Dulanska S, Remenec B, Matel L, Galanda D, Molnar A (2011) J Radioanal Nucl Chem 288:705-708
4. Ometakova J, Dulanska S, Matel L, Remenec B (2011) J Radioanal Nucl Chem 290(2):319-323
5. Izatt SR, Bruening RL, Krakowiak KE Izatt RM (2003) The selective separation of anions and cations in nuclear waste using commercially available molecular recognition technology products. In: Proceedings WM 03 Conf, Tuscon, 23-27 Feb 2003
6. TrisKem International, http://www.triskem-international.com/iso_album/ft_resine_dga_en_121219.pdf, cit. Accessed 28 August 2013

Primary author: Ms MILANOVIĆ, Ivana (Rudjer Boskovic Institute, Laboratory for Radioecology)

Co-author: Dr GRAHEK, Željko (Rudjer Boskovic Institute, Laboratory for Radioecology)

Presenter: Ms MILANOVIĆ, Ivana (Rudjer Boskovic Institute, Laboratory for Radioecology)

Session Classification: Poster Session - Radionuclides in the Environment, Radioecology

Track Classification: Radionuclides in the Environment, Radioecology